



**Frank Bacon Machinery Sales Co.**

**21251 Ryan Road**

**Warren, MI 48091**

**586-756-4280** [www.frankbacon.com](http://www.frankbacon.com) Email: [sales@frankbacon.com](mailto:sales@frankbacon.com)

Stock #	DESCRIPTION
13254-02	<p>4.5KIP Unholtz Dickie Model S452 Vibration System with Slip Table</p> <p>(1) USED 4.5KIP Unholtz-Dickie Vibration System, Model S452 Shaker (S/N 113) with SA15 Amplifier (S/N 3153) and Slip Table</p> <p style="text-align: center;"><b>Unholtz-Dickie Model SA15-S452 Vibration System</b></p> <p style="text-align: center;"><b>SPECIFICATIONS:</b></p> <p><b>Force Rating</b> <b>(Random):</b> 4,500 lbs peak <b>(Sine):</b> 4,500 lbs RMS <b>Stroke (Standard):</b> 2" peak to peak <b>Frequency Range (Hz):</b> DC to 3,500 Hz Max <b>Max Free Table Acceleration:</b> 90 g <b>Table Diameter:</b> 13.30" <b>Armature Weight:</b> 50 lbs <b>Total Weight (Lbs):</b> 7,500 lbs <b>Voltage Requirements:</b> 480V, 60Hz, 3PH, 38A</p> <p style="text-align: center;"><b>EQUIPPED WITH:</b></p> <ul style="list-style-type: none"><li>• Unholtz-Dickie S452 Shaker S/N 113</li><li>• Unholtz-Dickie SA30 Power Amplifier S/N 3153</li><li>• Unholtz-Dickie Field Supply FSE S/N 638</li><li>• Cooling Blower</li><li>• 36" x36" Slip Table (137 Lbs)</li><li>• Bullnose (23 Lbs)</li><li>• Miscellaneous Hoses, Cables &amp; Accessories</li></ul> <p style="text-align: center;"><b>FOB Warren, MI</b> <b>Cleaned, Serviced and Cycled</b> <b>30 Day Satisfaction Guaranteed Return Privilege</b> <b>90 Day Parts and Labor Warranty</b></p> <p>*Actual Pictures Prior to Cleaning &amp; Servicing</p>

**Guarantee:** Unless otherwise specified, every machine is offered with the standard **MDNA** (*MACHINERY DEALERS NATIONAL ASSOCIATION*) Return privilege to ensure your complete satisfaction. If the machine is un-satisfactory it may be returned to our warehouse, freight prepaid and in the original condition within 30 Days of shipment for a full refund less the cost of SPECIALIZED EQUIPMENT, NEW ITEMS AND RE-CERTIFICATION COSTS, When applicable. Care is taken to provide accurate specifications. However, Critical areas should be verified by Inspection.